



The Reichert Darkfield Quebec Colony Counter User Guide

13332500/13332600
13332700/13332800

Reichert[®]

Reichert Darkfield Quebec® Colony Counter User Guide

1.0 Introduction

The REICHERT DARKFIELD QUEBEC® Colony Counter provides efficient, accurate and convenient colony counting. Glarefree illumination over the entire culture plate assures clear, quick results. By touching the colony with the counting probe (Models 13332700/13332800), colonies are recorded on the digital display; they appear bright and are readily distinguished from other structures.

2.0 Unpacking & Assembly

Carefully remove the REICHERT DARKFIELD QUEBEC® Colony Counter and accessories from the packing box and inspect them to assure damage did not occur during shipping.

Place the instrument on a hard, flat surface.

A leg located on the bottom can be set at either a tilt angle or locked flat to the instrument base.

Similarly, the 1.5X magnification lens may be raised or lowered for user preference. A 1.5X auxiliary lens (Catalog No. 13333100) is available as an accessory and is designed to be placed over the existing lens to increase the total magnification to 3X.

The REICHERT DARKFIELD QUEBEC® Colony Counter (Model 13332700 & 13332800) includes a five-digit LCD counter with a reset button, a ground contact and counting probe (Figure 1). An optional push button counter also is available.

The counting probe (Models 13332700/13332800 only) is shipped with the needle tip reversed for protection from damage. To assemble the counting probe for use, loosen the knurled nut, remove the needle, place the blunt end in the chuck and retighten.

The ground contact has a magnetic base and is attached to the instrument case so the arm extends over the culture plate and touches the surface of the agar.

The push button counter plugs into the two-hole receptacle at the rear of the colony counter.

3.0 Testing the Instrument

Before using the REICHERT DARKFIELD QUEBEC® Colony Counter it should be tested. Plug the electrical cord into a grounded 120V/60 cycle AC outlet (Models 13332500/13332700) or 230V/50 cycle AC outlet (Models 13332600/13280000).

Turn the instrument on using the switch located on the rear panel. Touch the tip of the counting probe to the ground arm probe. A count should register on the LC.D. display, if the colony counter is operational.

To return the LCD display counter to "0" depress the reset button.

Explanation of Warning Symbols

 **CAUTION:** Replace with same type and rating fuse (IEC 60127-2/SS 3 standard).

 **CAUTION:** Risk of electric shock

 **CAUTION:** (Refer to accompanying documents)




Figure 1




Figure 2

4.0 Care of the Counting Probe

Care should be taken to keep the tip of the counting probe sharp. If the needle is blunt, small colonies will be difficult to count and accurately register.

The probe tips touching media and/or colonies should be cleaned as dictated by good laboratory procedures. Solutions of 70% isopropyl alcohol, iodine/alcohol, Lysol® Brand or formaldehyde may be used to clean the tip without harm.

5.0 Operating the Colony Counter

To count directly on the surface of the nutrient medium, place the uncovered culture plate on the culture plate holder. Loosen the two thumbscrews on the holder to center the culture plate on the counting plate.

Plug the ground contact and the counting probe into either electric receptacle on the back of the instrument. Position the ground contact with magnetic base near the culture plate so that the L-shaped tip makes contact with the surface of the agar.

Counts are automatically registered by simply touching each colony with the needle point of the counting probe (Figure 2). To prevent the possibility of spurious registration resulting from



Danger: The glass bulb may be hot. Do not attempt to change the bulb before it is completely cooled or without wearing adequate skin protection. !

Disconnect the instrument from its power source. Remove the screws on the face of the instrument. Swing the magnification lens aside and open the black front section attached to the lower hinge. Unscrew the burned out bulb and replace it. Reconnect the colony counter to its power source (Figure 3).




Figure 3

the "entry" and "exit" of the probe tip into the medium, a pulsed delay circuit will admit only one count over a short time period. This feature does not interfere with the maximum counting speed of a skilled operator.

6.0 Bulb Replacement

The REICHERT DARKFIELD QUEBEC® Colony Counter illumination source is a standard 40 watt soft white-type incandescent bulb. (50 watt for 230V operation)

7.0 Cleaning the Counting Plate

Remove the screw on the left side of the instrument face. Swing the magnification lens aside and open the black front section attached to the lower hinge. The counting plate is held in place with four screws attached to the back of the front panel. Loosen the screws and slip the plate out of the retainer.

Clean the counting plate with a solution of mild detergent and water. DO NOT USE AN ABRASIVE CLEANSER. It will damage the white grid on the counting plate.

Dry the counting plate with a lint-free cloth. Place the counting plate back into the retainer until it is firmly seated. Avoid leaving fingerprints on the surfaces, they may cause inaccurate readings. Ensure plate is installed with rulings facing bulb to avoid damage. Tighten the four screws and return the front section to the operating position. Replace the screw on the front of the instrument.

The REICHERT DARKFIELD QUEBEC® Colony Counter is designed to insure reliability, long life and minimal maintenance. Contact your local Reichert representative for additional operating or service information.

8.0 Fuses

The REICHERT DARKFIELD QUEBEC® Colony Counter has two fuses located in a fuse drawer next to the power switch on the rear panel. The instrument must be unplugged to allow access to the fuse drawer.



CAUTION: Use only the fuse specified on the data plate for replacement.

9.0 Specifications

The REICHERT DARKFIELD QUEBEC® Colony Counter bears the following marks: UL, cUL • Models 13332500 and 13332700 are rated 120VAC, 40 watts


- Models 13332600 and 13332800 are rated 230VAC, 50 watts

Operating Conditions


- Indoor use only
- RH 80% up to 31°C decreasing linearly to 50% RH at 40°C
- Pollution degree 2
- Altitude up to 2,000m
- Temperature 5°C to 40°C
- Main supply voltage $\pm 10\%$
- Installation CAT II

10.0 Probe Set Up

Models: 13332700/13332800




1. Loosen knurled collar and remove the standard probe.



2. Insert the probe handle into the hole on the magnetic base and tighten it in place.

3. Insert the "L" shaped probe into the front of the probe handle and tighten the knurled collar.



4. Plug both connector ends of the probes into either of the rear sockets.

5. Position the magnetic base onto the front cover of the Colony Counter so the "L" shaped probe extends over the edge of the petri dish and in contact with the Agar.

6. Use the opposite probe to touch the colony and advance the counter.

11.0 Accessories

Catalog No.	Description	Model
13332300	Counting Probes	13332700/13332800
13332400	Push Button Counter	13332500/13332600
13333100	1.5x Auxiliary Lens	ALL
13333800	Woffheugel Counting Plate	ALL



Reichert Analytical Instruments
Reichert, Inc.
3362 Walden Ave.
Depew, NY 14043
USA

Toll Free: 888-849-8955
Phone: 716-686-4500
Fax: 716-686-4555

email: ai@reichert.com
www.reichert.com

Reichert GmbH
Hubertusstrasse 4
D-82229 Seefeld
Germany
Tel: +49-8152-993530
ISO-9001 Certified

CE

Reichert[®]